Contribution ID: **2** Type: **talk**

List-mode TOF MLEM reconstruction for the total-body J-PET with a realistic system response matrix

Monday, 11 October 2021 09:20 (20 minutes)

We modify the time-of-flight maximum likelihood expectation maximisation (TOF MLEM) image reconstruction algorithm by an updated model for the system response matrix (SRM) of the total-body Jagiellonian PET (J-PET) scanners, which modular multi-layer geometry complicates SRM estimation and requires more computational power to calculate correction factors [1].

The elongated plastic scintillators of the J-PET, which use Compton scattering for the detection of positron-electron annihilation photons, imply the smooth dependence of SRM on the obliqueness angle θ. We thus represent it as a set of functions unique for each bin and acquired by a log-polynomial fit of the Monte Carlo simulated emissions of γ-photons on 2D transverse planes with different θ.

By utilising the GATE software [2], a NEMA IEC phantom [3] was simulated in a 140-cm long 24-module J-PET, comprised of 2 detector layers (inner radius 393 mm) and a layer of wavelength shifters [4]. The data collected from a 500-s long scan was post-smeared according to the assessed temporal (191 ps) and axial (5 mm) resolution. Only true coincidences were considered.

The updated SRM was employed for the list-mode TOF MLEM reconstruction. For the predefined NEMA IEC attenuation map, two versions of attenuation correction were applied: a conventional (integration over bins) and a simplified on-the-fly recalculation for each measurement, which improves performance and is less sensitive to boundary effects.

Compared to the reference list-mode TOF MLEM from the CASToR framework [5], a substantial improvement in the image quality and mean squared error with respect to ground truth were observed. The simplified attenuation correction proved to be a reliable alternative, producing outcomes similar or better than the conventional approach.

To summarise, the proposed analytical SRM model for the total-body J-PET proved to be superior to the reference method employed for crystal-based scanners. The modified TOF MLEM and attenuation correction do not require high computational power and can be extended to account for the non-collinearity, positron range and other factors.

[1] P. Moskal et al., "Simulating NEMA characteristics of the modular total-body J-PET scanner—an economic total-body PET from plastic scintillators," Phys. Med. Biol., vol. 66, no. 17, pp. 175015, Sep. 2021.

[2] S. Jan et al., "GATE: a simulation toolkit for PET and SPECT," Phys. Med. Biol., vol. 49, no. 19, pp. 4543-4561, Oct. 2004.

[3] Performance Measurements of Positron Emission Tomographs, NEMA NU 2-2012, 2013.

[4] J. Smyrski et al., "Measurement of gamma quantum interaction point in plastic scintillator with WLS strips," Nucl. Instrum. Methods Phys. Res. A, vol. 851, pp. 39-42, Apr. 2017.

[5] T. Merlin et al., "CASToR: a generic data organization and processing code framework for multi-modal and multi-dimensional tomographic reconstruction," Phys. Med. Biol., vol. 63, no. 18, pp. 185005, Sep. 2018.

Primary author: SHOPA, Roman (National Centre for Nuclear Research, Poland)

Presenter: SHOPA, Roman (National Centre for Nuclear Research, Poland)

Session Classification: Monday Morning Session